


Characteristics

Nitrile synthetic contains no natural rubber latex proteins. Soft, highly elastic formulation provides the maximum in comfort. Pair packaged, inner wrapper in ozone resistant multi-vac packaging.



**Exam Glove
Sterile Pairs**



NitriDerm®

**Powder-Free Nitrile
Series 106**

9 1/2" Length

PRODUCT DETAILS

SIZE	ITEM NO.	PACKAGING	DESCRIPTION
S	106100	50 Pairs/box, 4 boxes/case	Gloves, Exam, Nitrile, Sterile, Powder-Free, Pairs
M	106200	50 Pairs/box, 4 boxes/case	
L	106300	50 Pairs/box, 4 boxes/case	
XL	106350	50 Pairs/box, 4 boxes/case	

Product Attributes

- Low Modulus
- Non-Latex
- Textured Finish

Benefits

- Softer, More Comfortable Fit
- No Risk of Latex Allergens
- Improved Wet/Dry Grip

Product Solutions You Trust

NitriDerm®

STERILE PAIRS

Nitrile Examination Gloves

Powder-Free



NitriDerm® Sterile Pairs are manufactured in compliance with multiple international standards, including the following:

Designation	Standard
ASTM D6319	Standard Specification for Nitrile Examination Gloves for Medical Application
ASTM D5151	Standard Test Method for Detection of Holes in Medical Gloves
ASTM F1671	Standard Test Method for Resistance of Materials Used in Protective Clothing to Penetration by Blood-Borne Pathogens

Average Length	Average Palm Thickness	Average Finger Thickness
9.5 in ♦ 240 mm	4.0 mil ♦ 0.10 mm	4.5 mil ♦ 0.11 mm

Tensile Strength & Elongation	Before Aging	After Accelerated Aging
Tensile Strength (Mpa)	32	27
ASTM Requirement Min. (Mpa)	14	14
Elongation (%)	650	560
ASTM Requirement Min. (%)	500	400



Intertek

Innovative Healthcare Corporation
is certified to ISO 13485:2003 QMS
for medical devices.



4040 Norex Drive, Chaska MN 55318 phone:
952.448.4412 | fax: 952.448.6180
sales@smithengineeringinc.com
www.smithengineeringinc.com