




Specification Sheet

DermAssist® COATS™

Powder-Free Latex Exam Gloves • Colloidal Oatmeal Active Therapeutic System



CONTAINS COLLOIDAL OATMEAL USP, A SKIN PROTECTANT DRUG

Product Attributes

- Textured
- Low Modulus
- Low Protein
- Non-Detectable Residual Chemical Level

Benefits

- Improved Wet/Dry Grip
- Soft, More Comfortable Fit
- Reduced Chances of Latex Allergic Reactions
- Reduced Chance of Dermatitis and Type IV Reactions

DermAssist® COATS™ is manufactured in compliance with multiple international standards, including the following:

Designation	Standard
ASTM D3578	Standard Specification for Rubber Examination Gloves
ASTM D5151	Standard Test Method for Detection of Holes in Medical Gloves
ASTM F1671	Standard Test Method for Resistance of Materials Used in Protective Clothing to Penetration by Blood-Borne Pathogens
ASTM D5712	Standard Test Method for Analysis of Aqueous Extractable Protein in Natural Rubber

Average Length	Average Palm Thickness	Average Finger Thickness
9.5 in ♦ 240 mm	4.0 mil ♦ 0.10 mm	4.5 mil ♦ 0.12 mm

Tensile Strength & Elongation	Before Aging	After Accelerated Aging
Tensile Strength (Mpa)	28	24
ASTM Requirement Min. (Mpa)	18	14
Elongation (%)	850	820
ASTM Requirement Min. (%)	650	500



Intertek

Innovative Healthcare Corporation
is certified to ISO 13485:2003 QMS
for medical devices.



4040 Norex Drive, Chaska MN 55318 phone:
952.448.4412 | fax: 952.448.6180
sales@smithengineeringinc.com
www.smithengineeringinc.com