



NitriDerm® Scents **Great Grape™** Nitrile Synthetic Exam Gloves



- 100 Qty (By Weight)
- Single Use
- Ambidextrous
- Non-Sterile

This Product Is Made From 100% Nitrile Synthetic Polymer And Does Not Contain Natural Latex Proteins

NitriDerm® Scents Great Grape™ is manufactured in compliance with multiple international standards, including the following:

Designation	Standard
ASTM D6319	Standard Specification for Nitrile Examination Gloves for Medical Application
ASTM D5151	Standard Test Method for Detection of Holes in Medical Gloves
ASTM F1671	Standard Test Method for Resistance of Materials Used in Protective Clothing to Penetration by Blood-Borne Pathogens

Average Length	Average Palm Thickness	Average Finger Thickness
9.5 in ± 240 mm	3.2 mil ± 0.08 mm	4.0 mil ± 0.10 mm

Tensile Strength & Elongation	Before Aging	After Accelerated Aging
Tensile Strength (Mpa)	32	27
ASTM Requirement Min. (Mpa)	14	14
Elongation (%)	650	560
ASTM Requirement Min. (%)	500	400



Intertek

Innovative Healthcare Corporation
is certified to ISO 13485:2003 QMS
for medical devices.



4040 Norex Drive, Chaska MN 55318 phone:
952.448.4412 | fax: 952.448.6180
sales@smithengineeringinc.com
www.smithengineeringinc.com